



Measurement and simulation of all your process and temperature signals



A calibrator or standard is a for calibrating measurement

The advantage of this sort of instrument is that it can be used to calibrate an instrument without removing it from the process. It is also preferable to calibrate an instrument in its actual operating environment.

Not all quantities can be handled in this way, however. For the moment, on-site calibrators are mainly used for electrical signals and temperature.

The C.A 1621, C.A 1623 and C.A 1631 calibrators are ideal for industrial companies involved in developing and verifying measurement chains: whether you are an installer or a maintenance technician, whether you belong to the maintenance department or technical and general services, and whether you work in metallurgy, the steel industry, agri-food, hospitals or civil engineering, these instruments will meet your needs!

These calibrators offer two major functions: simulation and measurement.

SIMULATION.

The first function involves simulating a signal or generating a quantity with precisely-defined values. These quantities are applied upstream of the measurement chain.

MEASUREMENT

The second function, involving the measurement of the signal. If there is a panel meter at the end of the chain, you can compare its value with the one indicated by the calibrator. If there is no panel meter, the simulated input value is compared with the signal measured by the calibrator.

The whole range of calibrators (C.A 1621, C.A 1623 and C.A 1631)

is designed to accurately measure, generate or simulate all the low-level signals encountered in industry.

Because they are simple to use, they are ideal for on-site work. Due to the wide variety of sensors recognized, these instruments are suitable for all your applications.

Powered by batteries, these calibrators are stand-alone instruments offering a long battery life. When connected to the mains, they can be used continuously for as long as necessary!

a portable instrument nt instruments on site.

C.A 1621 temperature calibrator for thermocouple probes

The C.A 1621 is capable of measuring and simulating up to 8 types of thermocouple as well as a voltage in mV.

Specifications for measurement (input) / simulation (output):

	Input/Output range		Resolution		Accuracy
	- 10 mV to 100 mV		0.01 mV		± 0.025 % + 2 counts
	Function	Range	Resolution	Accuracy	Reference junction error
	Type J	-200 to +1,200°C	O.1°C	± (0.3°C + 10 μV)	± 0.3°C
	Type K	- 200 to + 1,370°C	0.1°C	± (0.3°C + 10 μV)	± 0.3°C
	Type T	- 200 to + 400°C	0.1°C	± (0.3°C + 10 μV)	± 0.3°C
4	Type E	- 200 to +950°C	0.1°C	± (0.3°C + 10 μV)	± 0.3°C
	Type R	- 20 to + 1,750°C	1°C	± (1°C + 10 μV)	± 0.3°C
	Type S	- 20 to + 1,750°C	1°C	± (1°C + 10 μV)	± 0.3°C
	Type B	+600 to +1,800°C	1°C	± (1°C + 10 μV)	± 0.3°C
	Type N	- 250 to + 1,300°C	0.1°C	± (0.3°C + 10 μV)	± 0.3°C



TYPE K - SK11

C.A 1623 calibrator for resistance temperature-probes

The C.A 1623 is capable of measuring and simulating up to 7 different types of probes and a resistor

Specifications for measurement (input) / simulation (output):

Range		easurement racy $\pm \Omega$	Simulation accuracy ±		Acceptable excitation in mA
0.00 Ω to 400	0.00 Ω	O.1 O.1	0.15 <mark>0.1</mark>		0.1 to 0.5 0.5 to 3.0
400.0 Ω to 1,5	500.0 Ω	0.5	0.5		0.05 to 0.8
1,500.0 Ω to 3	3,200.0 Ω	1 2	1		0.05 to 0.4
Mode	Range	4-wire input	— Accuracy in °C — 2-wire/3-wire input	Output	Acceptable excitation in mA
Pt10 385	- 200 to + 800 °C				0.1 to 3.0
Pt50 385	- 200 to + 800 °C	0.7	1.0	0.7	0.1 to 3,0
Pt100 385	- 200 to + 800 °C	0.33	0.5	0.33	0.1 to 3.0
Pt200 385	- 200 to + 250 °C +250 to + 630 °C	0.2 0.8	0.3 1.6	0.2 0.8	0.1 to 3.0
Pt500 385	- 200 to + 500 °C +500 to + 630 °C	0.3 0.9	0.4 0.3	0.6 0.4	0.05 to 3.0
Pt1000 385	- 200 to + 100 °C +100 to + 630 °C	0.2 0.5	0.2 <mark>0.2</mark>	0.4 0.2	0.1 to 3.0
Pt100 JIS	+200 to + 630 °C	0.2	0.5	0.3	0.1 to 3.0



C.A 1631 voltage/current process-signal calibrator

The C.A 1631 is a measuring instrument. It can be used for measurement or to deliver a DC current loop between 0 and 24 mA and a DC voltage between 0 and 20 V.

DC voltage input and output specifications

Calibre	Resolution	Accuracy ± (% of reading + counts)
100 mV	0.01 mV	0.02 % + 3
20 V	0.001 V	0.02 % + 3

Input impedance: 2 M Ω (rated value), < 100 pF Protection against overvoltage: 30 V Current delivered by rated voltage: 1 mA

DC current input and output specifications

Calibre	Resolution	Accuracy ± (% of reading + counts)
24 mA	0.001 mA	0.015 % + 3

Protection against overloads: 125 mA/250 V quick-response fuse

Display as percentage: 0% = 4 mA, 100% = 20 mA

Source mode: 1,000 Ω load at 20 mA for battery voltage \geq 6.8 V, (700Ω) at 20 mA for a battery voltage

between 5.8 and 6.8 V)

Simulation: external loop voltage condition: 24 V (rated value), 30 V maximum, 12 V minimum.

Loop voltage supply: 24 V ± 10 %



General specifications common to all three instruments

Unit	C.A 1621 and C.A 1623: °C or °F
Power supply	6 x 1.5 V
Dimension	205 x 97 x 45 mm
Weight	472 g (batteries included)
Mains power supply	Input: 100 V – 240 VAC, 50–60 Hz 1.8 A Output: 12 VDC, 2 A MAX



TO ORDER

C.A 1621 CALIBRATOR TEMPERATURE TK

Delivered with 2 thermocouple adapters, 6 AAA 1.5 V batteries,


C.A 1623 CALIBRATOR TEMPERATURE PT

Delivered with 2 pairs of test wires and crocodile clip, 6 AAA 1.5 V batteries,

C.A 1631 CALIBRATOR PROCESS

Delivered with 2 pairs of test wires and crocodile clip, 6 AAA 1.5 V batteries,

ACCESSORIES AND REPLACEMENT PARTS



FRANCE

Chauvin Arnoux

190, rue Championnet 75876 PARIS Cedex 18 Tel: +33 1 44 85 44 38 Fax: +33 1 46 27 95 59 export@chauvin-arnoux.fr www.chauvin-arnoux.fr

UNITED KINGDOM

Chauvin Arnoux Ltd

Unit 1 Nelson Ct, Flagship Sq, Shaw Cross Business Pk Dewsbury, West Yorkshire - WF12 7TH Tel: +44 1924 460 494

Fax: +44 1924 455 328 info@chauvin-arnoux.co.uk www.chauvin-arnoux.com

MIDDLE EAST

Chauvin Arnoux Middle East P.O. BOX 60-154

www.chauvin-arnoux.com

1241 2020 JAL EL DIB - LEBANON Tel: +961 1 890 425 Fax: +961 1 890 424 camie@chauvin-arnoux.com



1205 – Ed. 1 – 10/2011 - Non contractual document. Confirm specifications before ordering. - Athana